

Rigid Test Finger

(IEC 61032 - Test Probe 11)



Description of the device

The Rigid Test Finger is a precision test probe made according to Figure 7 (Fig. 7) of the IEC 61032 (Test probe 11) and is used to simulate a human finger. It is also used by the standards of CSA, IRAM, UL, IEC 60335, IRAM 4220-1 and in most of the rules involved in the verification of accessibility to live parts.

The Rigid Test Finger was made on stainless steel and Polyamide handle.

Technical characteristics

- Material: **Stainless Steel.**
- Handle: **Polyamide.**
- Surface finishing: **Polished.**
- According to: **IEC 61032 - Test Probe 11.**



Precision Tip

